

**Region I  
NMSS Licensee Event Report**

Licensee	AGILENT TECH				
Event Description					
License #	026 07-28762-	Docket #	03002988	MLER-RI	2002-070
Event Date	2002	Report Date	10/7/02		

**1. REPORTING REQUIREMENT**

<input type="checkbox"/>	10 CFR 20.2201 Theft or Loss	<input type="checkbox"/>	10 CFR 35.33 Misadministration
<input type="checkbox"/>	10 CFR 20.2203 30 Day Report	<input type="checkbox"/>	License Condition
<input type="checkbox"/>	10 CFR 30.50 Report		
<input checked="" type="checkbox"/>	Other 10 CFR 31.5		

**2. REGION I RESPONSE**

<input type="checkbox"/>	Immediate Site Inspection	Inspector/Date	
<input type="checkbox"/>	Special Inspection	Inspector/Date	
<input type="checkbox"/>	Telephone Inquiry	Inspector/Date	
<input type="checkbox"/>	Preliminary Report	<input type="checkbox"/>	Daily Report
<input checked="" type="checkbox"/>	Information Entered in RI Log	<input checked="" type="checkbox"/>	Review at next inspection
<input type="checkbox"/>	Report referred to: _____		

**3. REPORT EVALUATION**


<input checked="" type="checkbox"/>	Description of Event	<input type="checkbox"/>	Corrective Actions
<input checked="" type="checkbox"/>	Levels of RAM Involved	<input type="checkbox"/>	Calculations Adequate
<input type="checkbox"/>	Cause of Event	<input type="checkbox"/>	Additional Information Requested from Licensee

**4. SPECIAL INSTRUCTIONS OR COMMENTS**


**Public** ☒

**Non-Public** ☐

	<b>Initials/Date</b>		<b>Initials/Date</b>
Completed by:	BETSY ULLICH	Date	10/8/02
Reviewed by:	[Signature]	Date	10/9/2002

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1  
2002 OCT -7 PM 2:50

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

October 4, 2002


Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by customers from which we have obtained removable activity wipes in excess of 0.005  $\mu\text{Ci}$ . Cell serial # K4505 was returned from overseas and the others were domestic. The cells were evaluated by support engineering and their disposition is identified below. I have contacted the U.S. customers for further information and will notify the cell supplier of the high wipes.

Serial #	Model #	Activity ( $\mu\text{Ci}$ )	Disposition	Comments
C0732	18803	0.006	Scrapped	Cell 24 years old. Source dirty and dark.
S10736	19312	0.009	Scrapped	Cell 14 years old. Indication of excessive temperature. Source dark.
L5298	19235	2.489	Scrapped	Cell 13 years old. Cell is rusty and dark
K4505	G1533A	0.027	Refurbished	Cell is 4 years old. Source lightly darkened.

Please contact me at 302-633-8262 if there are any questions.

Thank you,

  
David S. Bennett  
Radiation Safety Officer